

## Section: Validierung

ID: 12

### Abstract-Title:

RAPID PROTOTYPING ANATOMIC MODELS FOR VALIDATION OF MEDICAL ROBOT SYSTEMS

### Authors:

*D. Malthan<sup>1</sup>, A. Schäfer<sup>1</sup>, F. Dammann<sup>2</sup>, E. Schwaderer<sup>2</sup>, R. Ciuman<sup>3</sup>, J. Rodriguez Jorge<sup>3</sup>, M.M. Maassen<sup>3</sup>*

<sup>1</sup> *Fraunhofer Institute Manufacturing Engineering and Automation IPA*

<sup>2</sup> *Department of Diagnostic Radiology, University Hospital of Tübingen*

<sup>3</sup> *Department of Otolaryngology-Head & Neck Surgery, University of Tübingen*

### Abstract-Text:

Supported by the DFG (German Research Foundation), Medical Navigation and Robotics Program (SPP 1124, MA 1458/2)

#### Purpose

Dedicated robot systems are more and more becoming a valuable tool in medical procedures. One of the most cited benefits resulting from the employment of complex mechatronical devices is the achievable precision of these systems in the micrometer range. On the other hand, the raw mechanical precision is of little significance, as limited resolution and distortions in CT imaging, segmentation and registration steps have a heavy impact on the overall tolerance chain.

In order to verify the precision of a medical robot system regarding the target application, a test scenario should be developed. The purposes of this test bed were to simulate realistic and reproducible surgical conditions and to allow a straight-forward interpretation of the test results.

#### Method

As a first step, three scenario settings were chosen: First, the milling of an implant bed for the insertion of the supporting electronic device of an auditory cochlea implant. By means of this scenario, the capability of a robot system to resect a fitting implant bed under protection of the dura mater should be evaluated. As a second scenario, the removal of the complete inner mastoid tissue was chosen. Here, the challenge was to remove all segmented tissue without violation of relevant risk structures like blood vessels or facial nerves. Third, a direct access to the middle ear in respect of the correct position and angle for the insertion of a bone anchored hearing aid (BAHA) should be milled. As starting point for all scenarios, a CT data set of the lateral skull base was recorded. Based upon this data set, the regions to be resected by the robot system as well as relevant risk structures were segmented manually. As a second step, STL (Stereolithography) surface geometry data was extracted. Using the STL data, rapid prototyping models with well-known characteristics were generated. Here, regions to be resected were coloured green, risk structures red, tolerance regions surrounding the resection regions by 1mm yellow and neutral tissues white.

The medical robot system was employed to mill the planned regions, the results were

inspected manually.

## Results

The robot assisted milling of the rapid prototyping models showed that the material of the employed RP procedure is comparable to the real bone tissue regarding the milling time, which is closely associated with the density of the resected tissue.

Also the reproducibility of the test bed performed very well, as different iterations of the procedure showed the same results. Thus, the test bed is suitable to identify optimal parameter settings for milling forces or feeds.

The false colour representation of the resection, neutral, risk and tolerance areas proved to be very useful for the estimation of the performance of the robot system.

## Conclusion

We introduce a generic model based upon a rapid prototyping false-colour test bed for the evaluation of medical robot systems performing milling tasks. Employing the test bed, mechatronic systems for the resection of bone tissues can be evaluated very close to realistic anatomical conditions. The described procedure is especially valuable for the optimization of registration and resection parameters, as a reproducible model is available at low price. Further steps, e.g. the geometric measurement of the resected volume, could increase the significance of the performed tests even further.

## Literature

Maassen M.M., Dammann F., Malthan D., Stallkamp J., Schwaderer E., Zenner H.P.: Development of a robot prototype with sensory feedback. 3rd Symposium of Middle Ear Mechanics in Research and Otology. Matsuyama, Ehime, Japan, 9th-12th July, 2003

Maassen M.M., Dammann F., Malthan D., Stallkamp J., Schwaderer E., Zenner H.P.: Erhöhung der Präzision von computer- und roboterassistierten Operationen an der lateralen Schädelbasis durch sensorische Rückkopplung. Jahrestagung der Deutschen Gesellschaft für Hals-, Nasen-, Ohrenkrankheiten, Kopf- und Halschirurgie. Dresden, 28. Mai bis 1. Juni 2003.

Dammann F., Schwaderer E., Seemann M.: Einsatz von Methoden der 3D-Visualisierung im Kopf-Hals-Bereich. Fortschr Röntgenstr 2003; 175: 86.

Malthan D., Stallkamp J., Wössner S., Dammann F., Zenner H.P., Maassen M.M.: Lasertriangulationsgesteuerte Sensorik und miniaturisierte Roboterkinematik können Implantationen an der lateralen Schädelbasis verbessern, 5. Oktober 2002, Leipzig, II Jahrestagung der Sektion Neuroendoskopie, Neuronavigation und intraoperative Bildgebung der Deutschen Gesellschaft für Neurochirurgie.